

Partial-wave microscopic spectroscopy detects subwavelength refractive index fluctuations: an application to cancer diagnosis

Hariharan Subramanian,¹ Prabhakar Pradhan,¹ Yang Liu,¹ Ilker R. Capoglu,¹ Jeremy D. Rogers,¹ Hemant K. Roy,² Randall E. Brand,² and Vadim Backman^{1,*}

¹Biomedical Engineering Department, Northwestern University, Evanston, Illinois 60208, USA

²Evanston Northwestern Healthcare, Evanston, Illinois 60208, USA

*Corresponding author: v-backman@northwestern.edu

Received October 3, 2008; revised December 12, 2008; accepted December 15, 2008;
posted January 5, 2009 (Doc. ID 102350); published February 12, 2009

Existing optical imaging techniques offer us powerful tools to directly visualize the cellular structure at the microscale; however, their capability of nanoscale sensitivity is restricted by the diffraction-limited resolution. We show that the mesoscopic light transport theory analysis of the spectra of partial waves propagating within a weakly disordered medium, such as biological cells [i.e., partial wave spectroscopy (PWS)] quantifies refractive index fluctuations at subdiffractional length scales. We validate this nanoscale sensitivity of PWS using experiments with nanostructured models. We also demonstrate the potential of this technique to detect nanoscale alterations in cells from patients with pancreatic cancer who are otherwise classified as normal by conventional microscopic histopathology. © 2009 Optical Society of America
OCIS codes: 300.6550, 290.1350.

Spectroscopy of elastic scattering is commonly used to probe tissue morphology [1]. However, the sensitivity of a light-scattering signal to refractive index fluctuations is significantly reduced when the size of the scattering structures falls below the wavelength (~ 500 nm). Recently, there has been significant interest in understanding biological systems at the nanoscale, which requires measurement of subwavelength refractive index fluctuations. According to the mesoscopic light transport theory [2–4], for an object that is weakly disordered and weakly scattering, it is indeed possible to probe refractive index fluctuations of any length scale, including those well below the wavelength [3,4], if one analyzes a signal generated by the multiple interference of one-dimensional (1D) propagating waves reflected from the refractive index fluctuations within the object. The enhanced sensitivity of 1D propagating waves to subwavelength correlation lengths of refractive index fluctuations l_c (i.e., $l_c < \text{wavelength } \lambda$) can be understood from the following consideration: while a three-dimensional scattering coefficient is $\sim (l_c/\lambda)^3$, and, thus, the contribution from small length scales is weighted down as l_c^2 , for 1D waves the scattering coefficient is $\sim (l_c/\lambda)$. The 1D propagating waves are one of many of a subset of waves (herein called 1D-partial waves) propagating within a scattering particle.

Recently, we reported an optical system [5] capable of isolating 1D-partial waves from different parts of a homogeneous scattering particle. A detailed description of the partial-wave spectroscopy (PWS) instrument is given elsewhere [5,6]. In brief, a broadband light with a spatial coherence length of $< 1 \mu\text{m}$ is focused onto the sample by a low-NA objective (Edmund Optics, NA of objective=0.4, NA of illumination=0.2, NA of collection=0.4). The illumination-beam diameter ($\sim 120 \mu\text{m}$) is much larger than biological cells ($\sim 8 \mu\text{m}$) and is well collimated within a cell located in the waist of the beam.

The resulting backscattered image is projected with a $60\times$ magnification onto the slit of an imaging spectrograph ($10 \mu\text{m}$ slit width) coupled with a CCD camera (Coolsnap HQ, Roperscientific, 1392×1040 $6.2 \mu\text{m}$ pixels) and mounted onto a motorized linear scanning stage (Zaber Technologies). The backscattering image is acquired by linearly scanning the slit of the spectrograph with a $10 \mu\text{m}$ step. The size of a pixel in the image plane (image pixel) is $6.2 \mu\text{m} \times 10 \mu\text{m}$, while the size of a pixel in the object plane (cell pixel) is $100 \text{ nm} \times 165 \text{ nm}$. At each scanning step x , the CCD camera records a matrix with one axis corresponding to λ and the other to the spatial position of the image y , resulting in a data cube (x, y, λ) . The system contains a flipper mirror that directs the image into a digital camera for quick visualization. The backscattering spectrum $I(\lambda; x, y)$ ($\lambda = 390\text{--}750$ nm, spectral resolution ~ 3 nm, spectral sampling ~ 0.25 nm; only the signal from $\lambda_1 = 500$ nm to $\lambda_2 = 670$ nm is analyzed owing to the low illumination and transmission efficacy of the instrument) is recorded for each cell pixel (x, y) . All spectra obtained from each cell pixel are normalized by the spectra of the incident light using mirror reflection. For each (x, y) , $I(\lambda)$ ($\equiv I(\lambda; x, y)$) is further processed to remove the high-frequency spectral noise using a sixth-order low-pass Butterworth filter with a normalized cutoff frequency of 0.08, and the variations in the lamp spectrum using a low-order polynomial $I_p(\lambda)$ fit to $I(\lambda)$. The normalized cutoff frequency (cutoff frequency/sampling frequency) was chosen such that it removes all oscillations below the spectral resolution of the spectrometer. The resulting spectrum $R(\lambda) = I(\lambda) - I_p(\lambda)$ is referred to as the fluctuating part of the reflection coefficient. It is important to realize that unlike traditional light scattering experiments, where a scattering signal is formed by all waves propagating within a scattering particle and interfer-

ing in the far field, the backscattering spectrum analyzed in PWS is formed by the subset of waves, in particular 1D-partial waves. The properties of both the object and the instrument facilitate the detection of 1D waves. While the low NA of the objective restricts the illumination and collection of light to a narrow cone, the weak refractive index fluctuations and a small radius of curvature of a cell spread on a glass slide reduces the probability of the interference among the adjacent 1D channel, which is further prevented by the low-coherence illumination. Thus, light interaction with a complex weakly disordered medium can be approximated as a combination of several independent parallel 1D channels with $R(\lambda)$ generated by the multiple interference of photons propagating in these 1D channels.

Typical $I(\lambda)$ and $I_p(\lambda)$ obtained from a particular cell pixel (in this case, a cell isolated from human pancreatic epithelium) spread on a glass slide is shown in Fig. 1(a). Figure 1(b) shows the corresponding spectrum $R(\lambda)$. To show the noise level, the spectrum is compared with the one obtained from a glass slide. As seen, the intensity fluctuations seen in the spectrum from the cell are above the noise floor. Absorption does not play a significant role in these spectral fluctuations, as the absorption coefficient μ_a is $\sim 1.5 \text{ cm}^{-1}$ [7] and the thickness of the cell is $\sim 4 \text{ }\mu\text{m}$. Similarly, the glass slide behind the cell does not contribute to the observed spectral fluctuations, as the spectrum of the reflection from the slide is expected to be flat. This brings us to the question of how $R(\lambda)$ is related to the properties of the object: thickness (L), average refractive index (n_0), the variance and the correlation length of refractive index fluctuations ($\langle \Delta n^2 \rangle$ and l_c). In the regime where the approximation of 1D independent channels is valid, $R(k)$ (where $k=2\pi/\lambda$ is the wavenumber) can be characterized using the 1D mesoscopic light transport theory [2–4]. Accordingly, the rms average of $R(k)$ can be written as $\langle R \rangle = L\xi^{-1}$, where ξ^{-1} is the scattering coefficient of a 1D channel. Although a complicated function of $\langle \Delta n^2 \rangle$ and l_c , ξ^{-1} can be simplified for $kl_c < 1$: $\xi^{-1} \propto 2k^2 L_d / n_0^2$ with $L_d = \langle \Delta n^2 \rangle l_c$. Following terminology used in condensed matter physics, L_d is referred to as the disorder strength. In the following discussion we consider the case of $kl_c < 1$ without the loss of generality [8]. If L and n_0 are known, L_d can be determined from $\langle R \rangle \approx 2k^2 L_d L / n_0^2$ for each cell pixel. In turn, L can be estimated from the autocorrelation function $C(\Delta k) = \langle R(k)R(k+\Delta k) \rangle / \langle R(k)R(k) \rangle$ [3,4] $\ln(C(\Delta k)) = -(\Delta k)^2 f(L_d) A L^\alpha$, where $A = \text{const}$ (in units of length) [4]. Function f and parameter α are numerically derived using finite-difference time-domain (FDTD) simulations. In a weakly disordered medium and in the absence of strong localization, f is a slowly varying function of L_d ($r^2=0.04$) and is approximated as a constant (i.e., $f \sim 1$); α arises owing to the finite spectral bandwidth ($\lambda_2 - \lambda_1$) of the spectrum that is being used to calculate $C(\Delta k)$ (in the limit of a very large bandwidth, $L(\lambda_2 - \lambda_1) / \lambda^2 \gg 1$, $\alpha=2$; for the bandwidth in our system $\alpha \sim 1$). $C(\Delta k)$ is calculated for the entire




Fig. 1. (Color online) (a) Normalized backscattering spectrum $I(\lambda)$ from a single pixel of a biological cell (solid blue) and a pure glass slide (dashed red) with the $I_p(\lambda)$ (black). (b) The corresponding spectrum $R(\lambda)$ obtained from a biological cell (solid blue) and from a pure glass slide (dashed red).

spectrum with a center wavelength $\lambda_{\text{mean}} = 585 \text{ nm}$ for each cell pixel. Therefore, knowing the experimentally obtained quantities $\langle R \rangle$ and $C(\Delta k)$ for a given cell pixel (x, y) and n_0 (assumed to be ~ 1.38), we can calculate L_d ,

$$L_d = B \frac{n_0^2}{2k^2} \langle R \rangle \frac{(\Delta k)^2}{-\ln(C(\Delta k))}, \quad (1)$$

where B is the calibration constant and $\ln(C(\Delta k)) / (\Delta k)^2$ is obtained by fitting a linear slope to $-\ln(C(\Delta k))$ versus $(\Delta k)^2$.

To confirm the hypothesis that the PWS scheme enables collecting 1D propagating waves and Eq. (1) is valid, we performed experiments on a series of nanostructured model media comprised of aggregated polystyrene nanospheres. The fabrication protocol is described in detail in [9]. In brief, the aqueous suspension of monodispersed polystyrene nanospheres (Duke Scientific, Inc.) of volume $\sim 50 \text{ }\mu\text{l}$ was uniformly smeared on a glass slide. The self-assembled lattice formed after 15 min of evaporation. We used models with L varying from 0.3 to 13 μm and nanosphere sizes 20, 40, 60, 80, 100, and 125 nm (standard deviations of sizes $\sim 10\%$). PWS measurements were obtained from 30 different combinations of L and nanosphere sizes. Since both l_c and Δn are known *a priori* in this model, i.e., l_c is approximately the size of a nanosphere and Δn is approximately the refractive index of a polystyrene, we were able to compare the actual L_d of the model with the L_d found from the experimental PWS data obtained using Eq. (1). As shown in Fig. 2, there was a good agreement ($r^2=0.97$) between the experimentally observed and the actual values of L_d . These experiments were also used to determine constant B in Eq. (1). Finally, this experiment demonstrated that the minimal length scale of refractive index fluctuations to which PWS is sensitive to is below 20 nm.

The nanoscale sensitivity of PWS could be critical to cell microscopy, where cells are regularly imaged to understand disease processes. Conventional cytology is not sensitive to changes in cell nanoarchitecture (e.g., ribosomes, membranes, nucleosomes, just to name a few cell structures with subdiffractional dimensions). At the same time, these are some of the




Fig. 2. Validation of the nanoscale sensitivity of L_d using experimental nanostructured model media.

most fundamental building blocks of the cell. We hypothesized that PWS can detect nanoarchitectural alterations in cells that are undetectable by cytology. We considered pancreatic cancer as a case in point. Pancreatic cancer is the fourth leading cause of cancer deaths in the U.S. with an overall five-year survival rate of $<5\%$. For diagnosis, pancreatic cells are extracted using fine needle aspirations and subjected to a cytopathological analysis. However, the sensitivity of cytology is low for mass lesions in symptomatic patients ($\sim 70\%$) and much lower for early lesions due in part to the relative rarity of frankly malignant-appearing cells that can be identified by cytology. We performed a pilot study on archival pancreatic cells (fixed with alcohol) obtained from 16 patients (seven normal and nine malignant). Six cases from these nine adenocarcinomas were cytologically diagnosed as cytologically normal. (We note that although fixation modifies the internal refractive index distribution compared to that of live cells, the intracellular morphology is expected to be maintained.) PWS measurements were obtained from three different cohorts of cells: cytologically normal cells from normal patients (N), cytologically malignant cells from cancer patients (C), and cytologically normal cells from cancer patients (CN). For each patient, ~ 40 cells were chosen at random. Typical bright field and PWS images (i.e., $L_d(x,y)$) obtained for these three cell types are shown in Figs. 3(a) and 3(b). As seen, the L_d image shows a clear difference between N and C cells. More importantly, L_d images are different between N and CN cells also. Further statistical analysis was performed using the two parameters that are obtained from the L_d maps, the mean and the standard deviation of the intracellular disorder strength ($L_d^{(c)}$ and $\sigma^{(c)}$). As shown in Figs. 3(c) and 3(d) both the $\langle L_d^{(c)} \rangle$ and $\langle \sigma^{(c)} \rangle$ ($\langle \rangle$ indicates the average taken over all the cells within a patient cohort) are highly significantly elevated in cancer patients compared to the control group ($P < 0.001$). A prediction rule developed using a linear regression model yielded 100% sensitivity and 100% specificity for cytologically normal patients versus cancer patients. Interestingly, the cytologically normal cells from cancer patients also had significantly elevated $\langle L_d^{(c)} \rangle$ and $\langle \sigma^{(c)} \rangle$ ($P < 0.001$) with 83% sensitivity and 100% specificity. We note that both $\langle L_d^{(c)} \rangle$ and $\langle \sigma^{(c)} \rangle$ vary within a patient cohort as indicated by the error bars in Figs. 3(c) and 3(d). However this variation ($\sim 50\%$ in normal population) is much smaller than the difference




Fig. 3. (Color online) (a) Representative bright field image and (b) corresponding pseudocolor L_d map recorded from three different cell types: N, C, and CN. (c) and (d) $\langle L_d^{(c)} \rangle$ and $\langle \sigma^{(c)} \rangle$ for different cell types. Error bars are the standard errors of the mean.

between the patient cohorts (which is $>100\%$). A higher L_d in cancer patients may be due to the increase in $\langle \Delta n^2 \rangle$ and/or l_c . Higher $\langle \Delta n^2 \rangle$ can be associated with the increased density of intracellular macromolecular complexes, while the change in l_c may be due to the macromolecular aggregation.

In summary, the backscattering spectrum from a weakly disordered medium contains spectral fluctuations that can be used to measure the disorder strength of the refractive index fluctuations within the scattering object. The disorder strength is sensitive to subwavelength nanoscale refractive index fluctuations. As an illustration of the potential capabilities of PWS, we showed that this technique may identify cancer cells by sensing microscopically undetectable alterations in cell architecture.

This work was supported in part by National Institutes of Health (NIH) grants R01 EB003682, R01 CA112315, and R01 CA128641, the V Foundation, and National Science Foundation (NSF) grant CBET-0733868.

References

1. K. J. Chalut, L. A. Kresty, J. W. Pyhtila, R. Nines, M. Baird, V. E. Steele, and A. Wax, *Cancer Epidemiol. Biomarkers Prev.* **16**, 223 (2007).
2. P. Pradhan and N. Kumar, *Phys. Rev. B* **50**, 9644 (1994).
3. S. B. Haley and P. Erdos, *Phys. Rev. B* **45**, 8572 (1992).
4. R. Rammal and B. Doucot, *J. Phys. (Paris)* **48**, 509 (1987).
5. Y. Liu, X. Li, Y. L. Kim, and V. Backman, *Opt. Lett.* **30**, 2445 (2005).
6. H. Subramanian, P. Pradhan, Y. Liu, I. R. Capoglu, X. Li, J. D. Rodgers, A. Heifetz, D. Kunte, H. K. Roy, A. Taflove, and V. Backman, *Proc. Natl. Acad. Sci. USA* **105**, 20124 (2008).
7. J. N. Qu, C. Macaulay, S. Lam, and B. Palcic, *Appl. Opt.* **33**, 7397 (1994).
8. M. Bartek, X. Wang, W. Wells, K. D. Paulsen, and B. W. Pogue, *J. Biomed. Opt.* **11**, 064007 (2006).
9. K. Chen, A. Taflove, Y. L. Kim, and V. Backman, *Appl. Phys. Lett.* **86**, 033101 (2005).